



Safe-T-Vue® • Product No. 7201-50 and 7201-200
QUALITY ASSURANCE LOT PERFORMANCE REPORT
 Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # **S10 041323**

Expiration Date: **2024-10-13**

Non - Chemistry

Mark ✓ for "Accepted" and X for "Not Accepted."

Box	Adh. Label	Foil Lid	Close Function	Fill Level
✓ Copy	✓ Copy	✓ Copy	✓ Formed	✓ Acceptable
✓ Tack	✓ Tack	✓ Position	✓ Locked/Stable	Underfill
✓ Clean	✓ Color	✓ Clean Peel	Hinge	Overfill
✓ Box Damage	Tray	Disk	✓ Elasticity	Color Label
	✓ Edges	✓ Appearance		✓ Color
	Color	✓ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change	
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C
Starting Test Temperatures	1 to 4 °C

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	Pass

Quantity Approved: 47,623

Performance Data: Red Spots - Random 9.3 °C; Red Filling in White Center 9.7 °C; Full Red Color at 10.0 °C.

Quality verification by: C Thomas Date 2023-07-18

Quality Management approval by: Approved Lot # S10 041323 ; Rose Daly Date 2023-07-18