

Safe-T-Vue® • Product No. 7201-50 and 7201-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 10 Irreversible Temperature Indicators

Lot # <u>S10 092523</u>

Expiration Date: **2025-03-25**

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох	Adh. Label	Foil Lid	Close Function	Fill Level
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill
√ Box Damage	Tray	Disk	√ Elasticity	Color Label
===0	√ Edges	√ Appearance		√ Color
	√ Color	√ Fit		

Test Protocol – Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change				
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C			
Starting Test Temperatures	1 to 4 °C			

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	125	2	0	Pass

Quantity Approved:	53,468	 -			
Performance Data: Red S	pots - Random <u>9.3 ° C;</u> Red	d Filling in W	hite <u>9.6° C</u> ; Full I	Red Color at 10.0 °	<mark>° С</mark> .
Quality verification by:			2023-10-24	^	
Quality Management appr	oval by: Approved Lot # _	510 092	523: Pou	She	Date <u>2023-10-3</u> 0
QCLF QAP0102-2 STV	QA Lot Performance Reportant Performance P	ort STV Re ration, 116 Th	v. 18 ne American Road,	Registry Doc. Morris Plains, NJ 07	