

Safe-T-Vue® • Product No. 7206-50 and 7206-200 QUALITY ASSURANCE LOT PERFORMANCE REPORT

Safe-T-Vue® 6 Irreversible Temperature Indicators

Lot # S6 111623

Expiration Date: **2025-05-16**

Non - Chemistry

Mark √ for "Accepted" and X for "Not Accepted."

Вох	Adh. Label	Foil Lid	Close Function	Fill Level	
√ Copy	√ Copy	√ Copy	√ Formed	√ Acceptable	
√ Tack	√ Tack	√ Position	√ Locked/Stable	Underfill	
√ Clean	√ Color	√ Clean Peel	Hinge	Overfill	
√ Box Damage	Tray	Disk	√ Elasticity	Color Label	
	√ Edges	√ Appearance		√ Color	
	√ Color	√ Fit			

Test Protocol - Finished Goods Quality Assurance

Test performed on random sampling of finished product per ANSI ASQZ1.4-2003, General Inspection Level 1, Double Sampling Plan for Attributes with AQL=1%. Electronic thermometers with probes are calibrated annually. Download Validation SOP, from www.zebra.com/safetvue.

Temperature Change and Corresponding Safe-T-Vue® Change		
Product Specification	Safe-T-Vue 6.0±0.3°C, Safe-T-Vue 10±0.3°C	
Starting Test Temperatures	1 to 4 °C	

Test Parameter	AQL	Sample Size	Accept Number	Test Result	Pass/Fail
Full Color Response Temperature	1%	80	1	0	Pass

Quantity Approved: 32,990	
Performance Data: Red Spots - Random 4.8 ° C; Red Filling in White Center 5.3 ° C; Full Red Color at 5.9 ° C.	
Quality verification by: Date Date	
Quality Management approval by: Approved Lot # <u>S6 111623</u> ; Role Daty Date <u>2023-11</u>	-3

QCLF QAP0102-2 STV QA Lot Performance Report STV Rev. 18 Registry Doc. #2046

Manufactured by: Temptime Corporation, 116 The American Road, Morris Plains, NJ 07950, USA

Tel: +1 (973) 630-6000 Fax: +1 (973) 630-6102 www.zebra.com